

# OIP development in practice: overcoming barriers to bioequivalence

Przemysław Reszka

CMC Regulatory Affairs Manager

BioBridges 2025, Prague











- Orally Inhaled Products (OIP) – regulatory framework
- In-vitro Therapeutic Equivalence (TE) – case study
- Updated guidelines – impact on pharmaceutical development and in-vitro approach
- Conclusions

# Orally Inhaled Products (OIP) – Regulatory Approach

Weight of Evidence (FDA) vs Stepwise Approach (EMA) ➡



**Generic DPI (1 strength) compared to REF** acc. OIP guideline CPMP/EWP/4151/00 Rev. 1 (still in force till 02.2026):

1. The product contains **the same active substance** (i.e. same salt...)

→ confirmed, the same salt as REF

2. The **pharmaceutical dosage form is identical** (e.g. ...DPI, etc.)

→ confirmed, inhalation powder hard capsules (DPI) with co-packaged single dose inhaler device

3. The active substance is in the solid state: any **differences in crystalline structure and/or polymorphic form** should not influence the dissolution characteristics, the performance of the product or the aerosol particle behaviour


→ single polymorphic form demonstrated, similar DD and APSD




4. Any **qualitative and/or quantitative differences in excipients** should not influence the performance of the product...

- Lactose comprises a major part of the formulation
- The same amounts of lactose as REF (lactose content declared in SmPC)
- Similar grade of lactose as in REF
- Similar XRPD, PSD, DD, APSD etc.


PSD:




XRPD:



SEM:



REF



TEST

## Case study - in-vitro comparison (3)

5. Any qualitative and/or quantitative **differences in excipients should not change the safety** profile of the product

→ No significant differences demonstrated


6. The **inhaled volume through the device** to enable a sufficient amount of active substance into the lungs should be similar (within +/- 15%)

→ Proposed device with same mechanism of de-agglomeration and equivalent resistance

7. **Handling** of the inhalation devices for the test and the reference products in order to release the required amount of the active substance should be similar

→ Handling similar to REF except for the way the capsule is punctured


→ No impact on DD






# Case study - in-vitro comparison (4)

8. The inhalation device has the same **resistance to airflow** (within +/- 15%)



→ Resistance to airflow „sameness” demonstrated

9. The **target delivered dose** should be similar (within +/- 15%)





→ Similar DD demonstrated across minimum (10th percentile), median and maximum (90th percentile) achievable flow rate in the target patient population



# Case study - in-vitro comparison (5)

## Flow rate (FR) dependency of dry powder inhalers





- No FR dependency demonstrated
- Point estimate of FPD of TEST within  $\pm 15\%$  of REF for each tested FR
- PK data (if applicable) can be obtained in healthy volunteers

# Case study - in-vitro comparison (6)

Data from the **complete particle size distribution profile** of individual and grouped stages of a validated multistage impactor...

APSD (middle flow rate)



Yu Liu, Heliyon 8 (2022) e10174

- The equivalence at all FR fulfilled (CI within 0.85-1.18), except mouthpiece, stage 7 and MOC
- These stages, in particular stage 7 and MOC are characterized by small deposition and high variability
- Deposition comparable on stages represented by particles of 1-6 µm (considered as clinically relevant)

# Case study - in-vitro comparison (7)

*The comparison should be performed per impactor stage **or justified group of stages...***

Group	Respective parts of human respiratory tract	T/R geometric mean ratio (middle flow rate)	CI lower	CI upper	Equivalence
Level 1	Oral cavity, Swallow fraction	1.106	1.097	1.113	YES
Level 2	Larynx, pharynx, primary bronchi	0.978	0.974	0.981	YES
Level 3	Bronchi, bronchioles	1.018	1.009	1.027	YES
Level 4	Alveoli	0.963	0.912	1.019	YES

- 4 groups were proposed representing major parts of human respiratory tract
- The equivalence at each group (at all flow rates) was demonstrated (CI within the 0.85-1.18)
- **No pre-specified protocol was available → Step 2 (PK) was initiated**



# Updated Quality guideline (Rev 1) – major changes for DPIs

<b>API</b>	<ul style="list-style-type: none"> <li>Complete description of micronisation process with in-process controls</li> </ul>
<b>Pharmaceutical development</b>	<ul style="list-style-type: none"> <li>3 batches of DP with at least 10 inhalers/units from each batch</li> <li>UDD and FPD over patient flow rate: FR 30-90 L/min or pressure drop 2-6 kPa</li> <li>Re-priming tested on worst-case scenarios</li> <li>Effect of environmental moisture: <ul style="list-style-type: none"> <li>25°C/70% RH or higher humidity specifically mentioned</li> <li>For capsule based DPIs lower humidity (e.g., 35% RH or 40% RH) to be tested</li> </ul> </li> <li>Simulation of dropping from EMA Q&amp;A</li> </ul>
<b>Manufacture</b>	<ul style="list-style-type: none"> <li>Conditioning time before batch release to be specified</li> <li>For combination products yield of the final assembling step to be reported</li> </ul>
<b>Excipients</b>	<ul style="list-style-type: none"> <li>Requirements for novel excipients defined</li> </ul>
<b>Specification - FPD</b>	<ul style="list-style-type: none"> <li>Validated alternative to multistage impactor (AIM – abbreviated impactor method)</li> <li>Specification range <math>\pm 25\%</math> from EMA Q&amp;A</li> </ul>
<b>Container Closure system</b>	<ul style="list-style-type: none"> <li>Pictures to be provided</li> <li>Reference to MDR</li> <li>Reference to <i>Guideline on the quality requirements for drug device combination products</i></li> </ul>
<b>Lifecycle management</b>	<ul style="list-style-type: none"> <li>Examples of significant changes provided</li> </ul>

- **Stepwise approach** *officially* required
- The **core in-vitro criteria remain unchanged** with some overlapping requirements being removed
- **APSD requirements specified:**
  - **Grouping limited to stages with low deposition** (<5% of DD)
  - Non-sized fractions (e.g. IP/Presep) can be one group
  - **At least four groups with defined cut-off** (non-overlapping, NMT 3 stages in 1 group)
  - Similarity concluded if the 90% CI for the observed ratio of the geometric means of TEST and REF within the acceptance limit of  $\pm 15\%$ , assuming log-normal distribution of data (85-118%)
  - **Not meeting the acceptance criteria at some stages/groups acceptable in exceptional cases**, but the number of samples should be sufficient to minimise the risk for Type II-error.
- **DPIs flow rate dependency (FRD):**
  - Resistance  $\pm 15\%$  vs REF: comparison at 3 pressure drops (2-6 kPa) or 3 FR
  - Resistance  $> 15\%$  vs REF: comparison at 3 pressure drops (2-6 kPa)
  - **Two types of graphs required** for FPD vs pressure drops/FR
  - **If dependency of TEST  $>$  REF: additional PK data with trained healthy volunteers or patients**

## Updated guidelines:

- Stepwise approach confirmed
- No revolutionary changes for DPIs – however each case requires separate evaluation
- DPIs previously developed by Polpharma meet the latest requirements

## Case study:

- Extensive development carried out with QTPP and CQA defined
- All studies performed acc. to the Quality guideline and OIP guideline
- Multiple REF batches tested and representative batches selected for in-vitro studies
- APSD comparable on group of stages but not on every single stage
- At the end of step 1 the Applicant concluded that TE needs to be supported by PK studies (step 2)
- **PK studies confirmed BE!**

→ Define your TE acceptance criteria in advance (pre-specified protocol) and stick to them!



# Thank You!

